KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

QUALITY CONTROL SYSTEMS LTD

ISO/IEC 17025:2017

Calibration Laboratory Number: **KENAS/CL/056**

PERMANENT ADDRESS OF LABORATORY

GATAKA PLAZA, OFF GATAKA ROAD, NGONG VIEW AVENUE

P.O Box 56871-00200

NAIROBI

Tel. No. +254 722692450

E-MAIL : info@qualitycsl.com

WEBSITE: www.qualitycsl.com


Laboratory Contact Person:

Janet Muriuki

Technical Signatories:

1. Loyford Kinoti – Pressure
2. John Wanjohi – Temperature

Expiry Date: 04-September-2029

Approved by: 

Date: 05-September-2025

KENAS ~~CEO~~/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – QUALITY CONTROL SYSTEMS LIMITED

1.0 Calibration Field: Pressure

Equipment to be calibrated	Measurand	Calibration method	Standard	Range	Calibration and Measurement Capability	Calibration Location
Bourdon tube and Vacuum Pressure gauges	Pressure	**Internal Method:** QCSP/OP/002 **Reference:** Guideline DKD-R 6-1:2014	Addiel Digital Reference Pressure Gauge **Traceability:** KEBS/MET/5/3/84/122	-0.1 MPa to 70 MPa	± 0.023MPa	Laboratory & Client sites

2.0 Calibration Field: Temperature

Equipment to be calibrated	Measurand	Calibration method	Standard	Range	Calibration and Measurement Capability	Calibration Location
Enclosed Chambers	Temperature	**Internal Method:** QCST/OP/004 Reference EURAMET cg. 20 V.5.0	Fluke Digital Thermometer **Traceability:** KEBS/MET/4/3/137/109	-100°C to 250°C	± 1.10°C	Laboratory & Client sites
Contact Thermometers		**Internal Method: QCST/OP/003** **Reference DKD-R 5-1**		-30°C to 400°C	± 0.85 °C	
Liquid Baths		**Internal Method: QCST/OP/007** **Reference DKD-R 5-4**		0°C to 250°C	± 1.10°C	
Sterilizers and Autoclaves		**Internal Method: QCST/OP/006** **Reference EURAMET cg.20 V.5.0**		70°C to 150°C	± 1.10°C	

*The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%.

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	05-September-2025	05-September-2025	04-September-2029

This schedule is issued subject to the terms and conditions of KENAS Accreditation.

Approved by:  Date: 05-September-2025

KENAS ~~CEO~~/ Authorized Representative

