KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

AFRICAN GAS AND OIL COMPANY LIMITED

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/063**

PERMANENT ADDRESS OF LABORATORY

AFRICAN GAS AND OIL COMPANY LIMITED

Plot No. 4838, Miritini

Off Old Mombasa – Nairobi Road,

P.O. BOX. 81881 – 80100

Mombasa, Kenya

Tel: +254 (0) 748484786

Email: ibtisam.said@agol.co.ke

Laboratory Contact Person:

1.Ibtisam Said

Expiry Date: 28-October-2028

Approved by: _____ Date: 15-July-2025

KENAS ~~CEO~~/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – AFRICAN GAS AND OIL COMPANY LIMITED

No	Testing Field	Test Object / Matrix	Test Parameter	Test Method	Type of Test	Measurement Range	Test Location
1	Chemistry	LPG	Copper Strip Corrosion	ASTM D1838	Physical	1A – 4C	AGOL Laboratory
			Oil Stain	ASTM D2158		Present / Absent	
			Residue	ASTM D2158		0 – 0.5 ml	
			Density Relative Density	ASTM D2163 / D2598 / D1657		0.5 -0.65 kg/l (500 -650 kg/m^3) 0.5 -0.65	
			Vapour Pressure	ASTM D2163 / D2598 / D1657		Max. 485 kPa	
			Hydrocarbon Composition	ASTM D2163		0.01 -100%	
		LPG Vapour	Ethyl mercaptan	ASTM D5305		Max. 50ppm	
			Hydrogen Sulphide	ASTM D4810		Max 5 ppm	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	29-Oct-2020	29-Oct-2020	28-Oct-2024
1st Re-accreditation	18-Jul-2024	29-Oct-2024	28-Oct-2028

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: _____ Date: 15-July-2025

KENAS CEO/ Authorized Representative