KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

NAKURU WATER AND SANITATION SERVICES COMPANY LIMITED

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/015**

PERMANENT ADDRESS OF LABORATORY

P. O. BOX 16314 - 20100
Nakuru, Kenya

Tel: +254 51 212269/214148

Email: info@nakuruwater.co.ke

Laboratory Contact Person:

John Ngatia Waweru

Expiry Date: 30-Nov-2028

Approved by: _____

~~KENAS CEO/~~ Authorized Representative

Date: 11-March-2025

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – NAWASSCO LIMITED

No	Testing Field	Test Object / Matrix	Test Parameter	Test Method	Type of Test	Measurement Range	Test Location
1.	Chemistry	Water and Wastewater	Total Nitrogen	NLTM 05-TN(AWWA/APHA:4500-E)	Spectrophotometry	0.1 to 3.0mg/l	NAWASSCO Chemistry Laboratory
			Total Phosphorous	NLTM 06-P(AWWA/APHA:4500-E)		0.02 to 1.30mg/l	
			Sodium	NLTM 13-Na(AWWA/APHA:3111-B)	Physical	0.002 to 1.0mg/l	
			Potassium	NLTM 13-K(AWWA/APHA:3111-B)		0.005 to 2.0mg/l	
			Calcium	NLTM 13-Ca(AWWA/APHA:3111-B)		0.003 to 20mg/l	
			Zinc	NLTM 14-Zn(AWWA/APHA:3111-B)		0.3 to 2mg/l	
			Copper	NLTM 14-Cu(AWWA/APHA:3111-B)		0.2 to 10mg/l	
			Lead	NLTM 14-Pb(AWWA/APHA:3111-B)		0.04 to 20mg/l	
			Nickel	NLTM 14-Ni(AWWA/APHA:3111-B)		0.5 to 10mg/l	
			Cadmium	NLTM 14-Cd(AWWA/APHA:3111-B)		0.05 to 2mg/l	
			Manganese	NLTM 14-Mn(AWWA/APHA:3111-B)		0.1 to 10mg/l	
			Iron	NLTM 14-Fe(AWWA/APHA:3111-B)		1.2 to 10mg/l	
			Sulphate	NLTM 15-SO_4(AWWA/APHA:3111-B)		10 to 75mg/l	
			Fluoride	NLTM16_F : AWWA/APHA:3111-B		0 – 3.0mg/l	
			Conductivity	NLTM 17-EC(WTW Multi 350i Meter)		0.1 to 1500µS/cm	

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

			Total Dissolved Solids	NLTM 17-TDS(WTW Multi 350i Meter)		0.1 to 1000mg/l	
			pH	NLTM 18-pH		0-14	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	01-DEC-2017	01-DEC-2017	30-NOV-2020
1st Re-accreditation	18-MAR-2021	18-MAR-2021	30-NOV-2024
2nd Re-accreditation	11-MAR-2025	11-MAR-2025	30-NOV-2028

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: _____ Date: 11-March-2025

KENAS CEO/ Authorized Representative