KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

MEDICAL LABORATORY SCIENCE COUNCIL OF NIGERIA NATIONAL LABORATORY EQUIPMENT CALIBRATION CENTER

ISO/IEC 17025:2017

Calibration Laboratory Number: **KENAS/CL/046**

PERMANENT ADDRESS OF LABORATORY

PLOT 1166 MOHAMMED N. UMAR LANE, DURUMI PHASE II GARKI

PMB 771 GARKI, ABUJA

ABUJA, NIGERIA

Tel. No. +234 902 956 9156

EMAIL: info-nlecc@mlscn.gov.ng

WEBSITE: www.mlscn.gov.ng


Laboratory Contact Person:

Eng. Olayinka Olabanji

Technical Signatories:

1. Olayinka Olabanji – Temperature/Mass
2. Mr. Kelechi Ogbonna – Volume/Mass
3. Ms. Abar Okoye Ifeoma – Rotary Speed

Expiry Date: 25-January-2027

Approved by: 

Date: 08-April-2025

KENAS ~~CEO~~/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – NATIONAL LABORATORY EQUIPMENT CALIBRATION CENTER

1.0 Calibration Field: Mass

Equipment to be calibrated	Measurand	Calibration method	Standard	Range	Calibration and Measurement Capability	Calibration Location
Non-Automatic Weighing Instruments	Mass	**Internal:** NALECC-TP-39 **References:** EURAMET cg-18 OIML R111-2004 OIML R076-1-2006	F1 Masses **Traceability:** EWML/CERT/A/250/013-01	1g to 2Kg	± 0.38g	Laboratory and Client Sites

2.0 Calibration Field: Temperature

Equipment to be calibrated	Measurand	Calibration method	Standard	Range	Calibration and Measurement Capability	Calibration Location
Contact Thermometers	Temperature	**Internal:** NALECC-TP-37 Reference BN-EN60068-3-7	Digital temperature Indicator, S/No.180913523 **Traceability:** TCSL-22-2067-06	30°C to 350°C	± 1.5°C	Laboratory and Client Sites
Controlled Heating Enclosed Chambers		**Internal:** NaLECC-TP-28 **Reference** DIN EN 60751	Digital temperature Indicator, S/No.170313771 **Traceability:** TCSL-22-2067-07	-40°C to 250°C	± 2 .5 °C	

3.0 Calibration Field: Rotary Speed

Equipment to be calibrated	Measurand	Calibration method	Standard	Range	Calibration and Measurement Capability	Calibration Location
Centrifuge	Rotational Speed	**Internal:** NaLECC-TP-29 **Reference** IEC 61010-2-020	Optical tachometers S/No. 160308450 **Traceability:** AIMS/CC/22/0266	100rpm to 15000rpm	±2rpm	Laboratory /Client Sites

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

4.0 Calibration Field: Volume


Equipment to be calibrated	Measurand	Calibration method	Standard	Range	Calibration and Measurement Capability	Calibration Location
Micropipettes	Volume	**Internal:** NaLECC-TP-34 **Reference** ISO 8566-6	F_2 masses **Traceability:** EWML/CERT/2203/034-01 Analytical Balances **Traceability:** TCSL-22-2067-09 TCSL-22-2067-08	10µl to 200µl 200µl to 1000µl	±0.5µl ±1.5µl	Laboratory /Client Sites

*The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%.

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	26-Jan-2023	26-Jan-2023	25-Jan-2027

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by:  Date: 08-April-2025

KENAS ~~CEO~~/ Authorized Representative

