KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

CSI INTERNATIONAL LIMITED

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/048**

PERMANENT ADDRESS OF LABORATORY

HOWSE & MCGEORGE CENTRE- GROUND FLOOR
FACTORY ST, BUILDING NO. 10, INDUSTRIAL AREA
OFF BUNYALA ROAD
P. O. BOX 47846- 00100
NAIROBI.

Tel: +254 (0) 712603528/254 (0) 705204610/254 (0) 723479111
Email: info@csiinternationalke.co.ke/ support@csiinternationalke.co.ke
Website: www.csiinternationalke.co.ke

Laboratory Contact Person:

Name of Contact Person — Goerge Oindo Achieng'

Expiry Date: 27-January-2029

Approved by: _____ Date: 27-January-2025

~~KENAS CEO~~/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – CSI INTERNATIONAL LIMITED

No	Testing Field	Type of Test	Test Method	Test Object / Matrix	Test Parameter	Measurement Range	Test Location
1.	Chemical	Gravimetric	CSI-TP-007	Water-Portable/ effluent	Total Suspended Solids	2-200 mg/L	CSI Chemical Laboratory
		Electrochemical	CSI-TP-012		Total dissolved solids	0-2000 mg/L	
			CSI-TP-004		Electrical Conductivity	0 – 200 mS	
			CSI-TP-017	Water	Chlorine	0-100 mg/L	
			HACH 8029		Fluoride	0.02 -2 mg/L	
			CSI-TP-02	Water & Soil	PH	2-12	
		Titrimetric	CSI-TP-07	Water	Chloride	0-100 mg/L	
		Spectrometric	CSI-TP-003	Water-Portable/effluent	Cu, Zn, Mn, Fe, Pb, Ca, Mg, Na, K	0-1000 ppm	
			HACH 8000		COD	0-2500 mg/L	
			HACH 8048	Water & Soil	Phosphates	0-500 mg/L	
			HACH 8051		Sulphates	0-500 mg/L	
			HACH 8039		Nitrates	0-500 mg/L	
2.	Microbiology	Enumeration	ISO 4832	Food	Coliforms	0 -300 cfu/plate	CSI Microbiology Laboratory
			ISO 4833-1		Total plate count	0 -300 cfu/plate	
			ISO 166492		*Escherichia coli*	0 -300 cfu/plate	
			ISO 6888-1		*Staphylococcus aureus*	0 -300 cfu/plate	
				Water		0 -100 cfu/plate	
		Enumeration	ISO 6222		Total viable count	0 -300 cfu/plate	
			ISO 9308-1		*Escherichia coli*	0 -200 cfu/plate	
					Coliforms	0 -200 cfu/plate	
			ISO 7899-2		*Enterococcus faecalis*	0 -200 cfu/plate	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	**28-JAN-2021**	**28-JAN-2021**	**27-JAN-2025**
1st Re-accreditation	**27-JAN-2025**	**28-JAN-2025**	**27-JAN-2029**

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: _____ Date: 27-January-2025

~~KENAS CEO~~/ Authorized Representative