



KENYA
ACCREDITATION
SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

KENYA REVENUE AUTHORITY INSPECTION AND TESTING CENTRE

ISO/IEC 17025:2017

Testing Laboratory Number: KENAS/TL/029

PERMANENT ADDRESS OF LABORATORY

Times Tower, Haile Selassie Avenue
P.O BOX 48240-00100
NAIROBI, KENYA

Tel: +254 (20) 2816199/310900 EXT 6199

Email: Paul.lkmat@kra.go.ke

Website: www.kra.go.ke

Laboratory Contact Person:

Paul Ikmat

Expiry Date: 07-July-2027

Approved by: _____

KENAS CEO/ Authorized Representative

Date: 10-April-2025

**SCHEDULE OF ACCREDITATION – KENYA REVENUE AUTHORITY INSPECTION AND TESTING
CENTRE**

No	Testing Field	Test Object / Matrix	Test Parameter	Test Method	Type of Test	Measurement Range	Test Location
1.	Chemistry	Alcoholic beverages/ Denatured Spirits	Ethanol Content	TM/16/01 Based on KS 05-610:2001	Distillation	10 – 100%	Food Laboratory
2.		Sugar	Polarization of Sugar	TM/16/05 Based on ICUMSA METHOD GS2/3-1	Polarimetry	90 - 100° Z	
3.		Fats and Oil	Saponification	TM/16/04 Based on AOCS method Cd 3-25	Titrimetric	180 – 255 mg KOH/g	
4.		Fats and Oil	Free Fatty Acids	TM/16/03 Based on AOCS method Ca 5a-40	Titrimetric	0.01 – 100% FFA	
5.		Fats and Oil	Iodine Value	TM/16/02 Based on AOCS method Cd 1b-92	Titrimetric	2 – 150 g/100g	
6.		Carbon Steel	Manganese Content	TM/16/06 Based on ASTM E350	Spectrophotometry	0.01 – 2%	Inorganic Laboratory
7.		Aluminium and Aluminium-based alloy	Aluminium	TM/16/07 Based on ASTM E34	Spectrophotometry	0.3 – 100 mg/L	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	08-JUL-2016	08-JUL-2016	07-JULY-2019
1 st Re-accreditation	20-FEB-2020	20-FEB-2020	07-JULY-2023
2 nd Re-accreditation	08-JUNE-2023	08-JULY-2023	07-JULY-2027

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: 

Date: 10-April-2025

KENAS ~~CEO~~ Authorized Representative

