KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

VIVO ENERGY UGANDA

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/062**

PERMANENT ADDRESS OF CAB

PLOT 9/11, 7TH STREET, INDUSTRIAL AREA

P.O. BOX 7082, KAMPALA UGANDA.

Telephone (+256) 312210010/ 772754028

Email: Kennedy.Ggwanga@vivoenergy.com, Josephine.Nakato@vivoenergy.com

Laboratory Contact Person:

Ggwanga Kennedy

Nakato Josephine

Expiry Date: 09-March-2028

Approved by: _____ Date: 13-October-2025

KENAS CEO/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – VIVO ENERGY UGANDA TESTING LABORATORY

No.	Testing Field	Test Object/ Matrix	Test Parameter	Test Method	Type of Test	Measurement Range	Test Location
1.	Chemistry	Gasoline and Diesel	Distillation	ASTM D86-18	Volatility	400°C Max	VEU-Testing Laboratory
				ASTM D93A-18 Method A		19°C - 400°C	
		Diesel	Flash Point	ASTM D1298-12b	Gravimetric	0.775 - 0.840 Kg/m^3 @15°C	
		Gasoline and Diesel	Density	ASTM D4052-18a		0.68 – 0.97 Kg/m^3	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	10-Mar-2020	10-Mar-2020	09-Mar-2024
1st Re-accreditation	21-Dec-2023	10-Mar-2024	09-Mar-2028

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: _____ Date: 13-October-2025

~~KENAS CEO~~/ Authorized Representative