



PL-35-01

**Proficiency Testing and
Interlaboratory Comparison Policy**



Table of Contents

1	Background Information	2
1.1	Purpose	2
1.2	Scope	2
1.3	Legal Basis	2
2	Terms and Definitions	2
2.1	Acronyms and Abbreviations	2
2.2	Definition of Terms	3
3	Policy Statement.....	3
4	Associated Documents	6
5	Revision/Amendment Records.....	6



1 Background Information

1.1 Purpose

Assuring the quality of test and calibration results requires laboratories to have quality control procedures for monitoring the validity of test and calibrations undertaken. The monitoring may include the participation in inter-laboratory comparison or proficiency testing programmes or participation in external Quality Assurance programmes. Where these do not exist, a justification has to be provided on the absence of or reason for non- participation while other means that include regular use of certified reference material or replicate test or calibrations using the same or different methods are employed

Kenya Accreditation Service (KENAS) considers the participation of laboratories in external proficiency testing / inter-laboratory comparisons an important mechanism for monitoring the integrity of test / calibration results. This statement sets out the policy of the Kenya Accreditation Service (KENAS) with respect to participation in proficiency testing / inter-laboratory comparisons by Testing and Calibration Laboratories.

1.2 Scope

This policy applies to the Laboratory accreditation schemes and Inspection schemes where testing results are used as part of the inspection activity.

1.3 Legal Basis

ISO/IEC 17011 provides for requirements for accreditation bodies accrediting conformity assessment bodies. ILAC policy, ILAC P9 for participation in proficiency testing activities.

2 Terms and Definitions

For the purposes of this document, the terms and definitions given in ISO/IEC 17000 and the KENAS Act 2019 and the following apply.

2.1 Acronyms and Abbreviations

KENAS	Kenya Accreditation Service
PT	Proficiency Testing
ILC	Inter-Laboratory Comparison

2.2 Definition of Terms

2.2.1 Proficiency Testing

The evaluation of a calibration or testing performance of a laboratory against pre-established criteria by means of inter-laboratory comparison.

2.2.2 Interlaboratory Comparison

The organization, performance and evaluation of measurements or tests on the same or similar calibration/test items by two or more laboratories or inspection bodies in accordance with predetermined conditions.

3 Policy Statement

- 3.1 This policy applies to all KENAS accredited laboratories and those laboratories seeking accreditation participate in proficiency testing / inter-laboratory comparisons and achieve satisfactory outcomes where such schemes are available and are relevant to their scope of accreditation or scope being sort for accreditation. The same also holds for inspection bodies that perform testing activities which inform inspection outcomes.
- 3.2 Accredited laboratories must participate, as a minimum, in at least one satisfactory PT scheme for each parameter within the laboratory's scope of accreditation, between periods of re-assessment except where regulatory requirements require a higher frequency of PT participation.
- 3.3 The plan for PT participation needs to be clearly drawn with a clear rationale on the frequency and level of participation and shall be regularly reviewed in response to changes in staffing, methodology, instrumentation and any other changes that could have an impact on the validity of results. Laboratories must be prepared to justify their policy and approach to both frequency of participation and any non-participation in readily available PTs that are appropriate.
- 3.4 Calibration laboratories shall participate in ILC for the items on their schedule or proposed schedule of accreditation, including specific instruments or measurement devices where these are separately listed. These items, instruments or measuring devices shall be addressed in the PT activity plan / Protocol.
- 3.5 The PT / ILC activity plan should cover;

- a) Participants
 - b) The PT / ILC Scheme
 - c) The proposed measurement artefact or instrument / or proposed test method
 - d) The measurement / test parameters
 - e) Establishment of the reference values
 - f) Acceptance Criteria
 - g) Responsibility for issue of the PT/ILC report.
- 3.6 Accredited laboratories must participate in, at a minimum, one satisfactory inter-laboratory comparison covering the scope and sub-scope of accreditation between periods of re-assessment except where regulatory requirements require a higher frequency of PT participation.
- 3.7 Laboratories preparing for accreditation are required to participate and achieve satisfactory performance in a proficiency testing or inter-laboratory comparison where such schemes are available and are relevant to their scope of accreditation before a recommendation for accreditation can be considered.
- 3.8 Where no appropriate proficiency testing or inter-laboratory comparison is available, the laboratories are required to justify the lack of participation and provide for how they demonstrate the validity of their tests and calibrations by other means such as use of certified reference materials, replicate tests or calibrations using the same or different method for verification.
- 3.9 The KENAS assessment team will review at each assessment and surveillance visit the laboratory's performance in proficiency testing / inter-laboratory comparisons and the plans thereof.
- 3.10 Laboratories are required to have appropriate acceptance criteria and procedures for investigating the cause of problems and for implementing corrective actions when the acceptance criteria are not met. A written record of these activities must be maintained. The laboratory must ensure that it does not claim accreditation for any tests that could be affected by the events that caused "out of specification" proficiency testing / inter-laboratory comparison results until it is satisfied that the investigation into the anomalous result has fully resolved the issue. In the event that the laboratory establishes that test results are compromised, they shall inform KENAS and seek voluntary suspension for the test(s) in question.

- 3.11 If at any time the laboratory's performance in proficiency testing / inter-laboratory comparison in the opinion of KENAS, casts doubt on the integrity of test results. KENAS may suspend the relevant tests from the laboratory's scope of accreditation. The laboratory will be required to provide KENAS with written evidence that the problem has been identified and satisfactorily rectified (which may include demonstrated satisfactory performance in subsequent proficiency testing / inter-laboratory comparisons) before re-instatement of accreditation can be considered.
- 3.12 KENAS may specify participation in a particular scheme/exercise where it is deemed necessary to demonstrate technical competence. Where participation in an externally run scheme is a mandatory requirement, this shall be stated in an appropriate KENAS.
- 3.13 Where justification show that participation in externally run PT / ILC programs is not feasible, and or that no formal PT/ILC programs are available, the laboratories shall indicate other inter-laboratory comparison activities in which they intend to participate. These may include activities arranged by themselves (bilateral or multilateral inter-laboratory comparisons etc.) with other laboratories in order to satisfy this requirement.
- 3.14 The PT/ILC reports shall be clear and comprehensive and include at least the following information;
- a) Participants
 - b) Measurement/ test protocol
 - c) Identification of measurement standard /Artefact
 - d) Measurement / test results
 - e) Reference values and how these were established
 - f) Evaluation of measurement/test results
 - g) Performance of individual participants
 - h) Acceptance criteria
- 3.15 The KENAS website (www.kenas.go.ke) provides a link to known proficiency testing service providers. It is the responsibility of the Laboratory to identify that which is suitable for their need

4 Associated Documents

Ref	Document Identifier	Document Title
1.	ISO/IEC 17011	Requirements for accreditation bodies accrediting conformity assessment bodies.
2.	ISO/IEC 17025	General requirements for the competence of Testing and Calibration.
3.	ISO 15189	Medical laboratories – Requirements for quality and competence.
4.	ILAC P9	ILAC Policy for participation in Proficiency Testing Activities.

5 Revision/Amendment Records

Date	Ver	Revised By	Reason for Revision
10/10/2022	01	RSQ	Newly developed/formatted document